

# DISSOLVED OXYGEN TRANSMITTER

Model : TR-DOT1A4

ISO-9001, CE, IEC1010



**Lutron**

**LUTRON ELECTRONIC**

**NDTKALA<sup>®</sup>.CO**  
Non-Destructive Test (NDT) Equipment

تامین تجهیزات و مواد مصرفی تست های غیر مخرب ایران  
۰۲۱-۷۱۰۵۳۸۸۸ | ۰۹۱۲۰۲۴۶۲۷۰

**The Art of Measurement**

# DISSOLVED OXYGEN TRANSMITTER

Model : TR-DOT1A4

<b>Measure Range</b>	<p><i>Range 1 :</i> O<sup>2</sup> CAL.</p> <p><i>Range 2 :</i> DO : 0 -20 mg/L</p>
<b>Output</b>	<p>4 - 20 mA</p> <p><i>Range 1 ( CAL ) :</i> 20.9 % O<sup>2</sup> = 12 mA</p> <p><i>Range 2 ( DO ) :</i> 0 mg/L = 4 mA 10 mg/L = 12 mA 20 mg/L = 20 mA</p>
<b>External Adjustment</b>	ZERO VR SPAN VR
<b>Power Supply</b>	90 - 260 ACV 50 Hz/60 Hz
<b>Operation Temperature</b>	0 - 50 °C 32 - 122 °F
<b>Operation Humidity</b>	Less than 80 %RH
<b>Accessory Included</b>	Dissolved Oxygen Probe Probe-filling electrolyte, OXEL-03 Spare diaphragms, OXDP-02
<b>Optional Accessories</b>	Probe-filling electrolyte, OXEL-03 Spare diaphragms, OXDP-02
<b>Mounting</b>	Din rail or Wall
<b>Case</b>	ABS plastic
<b>Others</b>	Separate probe with plug

\* Appearance and specifications listed in this brochure are subject to change without notice.